

Corrigendum

Tender for supply of incubator is uploaded on eprocurement website vide tender id 2017_DARE_163418_2. Specifications for incubator should be read as follows

Specification for Incubator

- Capacity – 80-100 liters
- AISI 304 stainless steel chamber
- Temperature controller - PID microprocessor temperature controller, temperature set by control keypad to 0.1°C, digital LED
- Temperature range – Ambient +3°C to 80°C
- Temperature accuracy - $\pm 0.1^\circ\text{C}$
- Temperature Uniformity - $\pm 0.5^\circ\text{C}$ at 37°C
- Temperature distribution by forced air circulation
- No. of adjustable shelves – 2; **with inner glass door**
- Thermostat for over temperature protection and **anti-microbial coating for external surfaces**
- Timer for continuous operation or 1 min to 99 hours
- Quotation must include printed brochure with all the technical details and valid authorized dealership certificate
- ISO or internationally accepted rating
